## **Latex** Gloves



PRODUCT INFORMATION					
MATERIAL	Latex*				
COLOUR	Natural				
TYPE	Ambidextrous, non-sterile, single-use				
INTERIOR	Powder-free and polymer coated				
EXTERIOR	Fully textured				
SIZES Measure widest part of the knuckles.	<ul> <li>XS = 75mm +/-5</li> <li>S = 85mm +/-5</li> <li>M = 95mm +/-5</li> <li>L = 106mm +/-5</li> <li>XL = 116mm +/-5</li> </ul>				
COUNTRY OF ORIGIN	Malaysia				
STORAGE	Store in original packaging in a cool, dry and well ventilated area, away from dust, direct sunlight, moisture, x-ray and excessive heat above 100°F (37°C)				

PHYSICAL PROPERTIES					
AQL	1.5				
GLOVE WEIGHT	6.0grams (medium)				
GLOVE THICKNESS	5mil				
GLOVE LENGTH	240mm				
	BEFORE AGING	AFTER AGING			
TENSILE STRENGTH (MPA)	min. 18	min. 14			
ULTIMATE ELONGATION	min. 650% min. 500%				





QUALITY STANDARDS				
FDA STATUS	(21 CFR 177) compliant for food handling (510k) cleared for medical use			
AUDIT STANDARDS	Manufactured in an ISO 9001:2015 and an ISO 13485:2016 facility Manufactured in a Certified WRAP Facility			
TEST STANDARDS	ASTM D5712-05 - below 50 μG/G water extractable protein ASTM D3578 and EN 455 EN ISO 374-1:2016/Type C EN ISO 374-5:2016 Resistance to Bacteria, Fungi & Virus			

PACKAGING & ORDERING INFORMATION						
CODE	SIZE	PURCHASE UNIT	CARTON DIMENSIONS (LxWxH)	CARTON WEIGHT	CUBIC METRE	
109010	XS		36.5 x 25 x 24.5cm	7.3kg	0.02m³	
109020	S	1 carton of 1000 Gloves (100/box x 10)				
109030	М					
109040	L					
109050	XL					

RESISTANCE OF GLOVES TO PERMEATION BY CHEMICALS								
CHEMICAL			EN ISO 374-1:2016 PERFORMANCE LEVEL		EN 374-4:2013 MEAN DEGRADATION / %			
Diethylamine (G)			0		7.2			
Sodium Hydroxide 40% (K)			6		-14.9			
Sulphuric Acid 96% (L)	0		100.0					
Hydrogen Peroxide 30% (P)			2		-15.6			
Formaldehyde 37% (T)			1		-22.4			
EN ISO 374-1:2016 - permeation levels are based on breakthrough times as follows:								
Performance Level:	1	2	3	4	5	6		
Minimum breakthrough time (Min):	>10	>30	>60	>120	>240	>480		

**EN 374-4:2013 -** Degradation results indicate the change in puncture resistance of the gloves after exposure to the challenge chemical

Safety gloves to protect against chemicals are classified according to their permeation time (time taken for the chemical to penetrate the glove) and number of chemicals tested:

- Type A at least 30min each for at least 6 test chemicals
- Type B at least 30min each for at least 3 test chemicals
- Type C at least 10min each for at least 1 test chemicals

EN ISO 374-5:2016 - Resistance to Bacteria and Fungi = Pass, Resistance to Virus = Pass

## **MANDATORY STATEMENTS EN ISO 374-1:2016**

"This information does not reflect the actual duration of protection in the workplace and the differentiation between mixtures and pure chemicals."

"The chemical resistance has been assessed under laboratory conditions from samples taken from the palm only (except in cases where the glove is equal to or over 400mm - where the cuff is tested also) and relates only to the chemical tested. It can be different if the chemical is used in a mixture."

"It is recommended to check that the gloves are suitable for the intended used because the conditions at the workplace may differ from the type depending on temperature, abrasion and degradation."

"When used, protective gloves may provide less resistance to the dangerous chemical due to changes in physical properties. Movements, snagging, rubbing, degradation caused by the chemical contact etc. may reduce the actual use time significantly. For corrosive chemicals, degradation can be the most important factor to consider in selection of chemical resistant gloves."

"The penetration resistance has been assessed under laboratory conditions and relates to the tested specimen."



Contact us today to receive samples or for more information on this product.



